

TDS Code: D8189008



青岛伦敦杜蕾斯有限公司
QINGDAO LONDON DUREX CO., LTD.

地址: 青岛高新技术产业开发区香东路上马段 1-13 号
Address: No. 1-13 Shangma Part, Ao-dong Road High-tech Industrial Development Zone, Qingdao, China
邮编 Post Code: 266114 电话 Tel: (0086) 0532 66966193 传真 Fax: (0086) 0532 87838708

产品合格证明 Certificate of Conformance

合格证编号 No. of Certificate **COC354324**

兹证明该批产品符合 GB/T 7544 / ISO 4074 标准<<天然橡胶胶乳男用避孕套技术要求与试验方法>>
This certificate certifies that this batch of product conforms with the standard GB/T 7544 / ISO 4074
Natural Rubber Latex Male Condoms – Requirements and Test Methods.

产品名称 Description of Product: 天然胶乳橡胶避孕套 Natural Latex Rubber Condom

产品商标 Trade Mark: 杜蕾斯® Durex® 或 or 杜蕾斯™ Durex™

品名 Brand: 活力装 Jeans

产品批号 Lot No.: 1002548735

产品识别码 ID No.: N/A

生产日期 Manufacture Date: 2024-06

失效日期 Expiry Date: 2029-05

产品代码 FG Code: 3063162

包装数量 Pack Size: 12's

内包产品代码 Foiled Condom Code: 3155286

报告人 Reported By: 李吉非

日期 Date: 04/07/2022

批准人 Approved By: 李吉芳

日期 Date: 04/07/2022



参阅号 Ref: QSF 8.2-5.0201 签发 Issued: 10/2022 页 Page: 1 of 1 日期 Date: 10/2022





青岛伦敦杜蕾斯有限公司 QINGDAO LONDON DUREX CO., LTD.

题目Subject: 分析证明 Certificate Of Analysis

地址: 青岛高新技术产业开发区香东路上马段1-13号

Address No.1-13 Shangma Part, Aodong Road High-tech Industrial Development Zone, Qingdao,

邮编Post Code: 266114 电话Tel: (0086) 0532 66966193 传真Fax: (0086) 0532 87838708

品名 Brand: 活力装Jeans

颜色 Color: 粉红Coral

内包数量 Quantity Foiled (只 PCS): 297936

产品类型 Type: 52mm

润滑液类型 Lubricant Type: 硅油Silicone oil, 香草Vanilla

包装数量 Quantity Packed (托 CAR): 2069

批号 Lot No.: 1002548735

生产日期 MFG. Date: 2024-06

样品收到日期 Sample Arrival Date: 03/07/2024


包装规格 Pack Size: 12's

失效日期 EXP. Date: 2029-05

检测时间 Tested Date: 03/07/2024

检验标准Tested According To: GB/T 7544 / ISO 4074

技术要求 Requirements	抽样量 (只) Sample Size (PCS)	AQL	限值 Limits	不合格水平 Reject Level	检验结果Test Results			
					范围 Range	平均值 Average	不合格数量 NO. Failed	合格/不合格 Passed /Failed
尺寸 Dimmension:								
长度Length (毫米mm)	13	N/A	160 Minimum.	1	177-182	180	0	合格Passed
宽度Width (毫米mm)	13		标称 STD ± 2	1	51.5-52.5	52.0	0	合格Passed
厚度Thickness(毫米mm)	13		***根据不同规格 Individual Spec	1	0.053-0.055	0.054	0	合格Passed
爆破检验 Burst Test:								
老化前Initial 体积Volume (升l)	200 []	1.5	**19.0 Minimum. []	8 []	27.5-49.0	40.5	0	合格Passed
	315 [v]		22.0 Minimum. [v]	11 [v]				
	500 []		*25.0 Minimum. []	15 []				
其它Other ____			****					
压强Pressure (千帕 kPa)			1.0 Minimum.		1.15-1.75	1.55		
针孔 Freedom From Holes								
	315 [v]	0.25		3 [v]	-	-	0	合格Passed
	500 []			4 []	-	-		
其它Other ____					-	-		
可见缺陷 Visible Defect								
	315 [v]	0.40		4 [v]	-	-	0	合格Passed
	500 []			6 []	-	-		
其它Other ____					-	-		
独立包装可见开口 Individual Containers With Visibly Open Seals								
	315 [v]	0.40		4 [v]	-	-	0	合格Passed
	500 []			6 []	-	-		
其它Other ____					-	-		
包装密封性 Package Integrity								
	32 [v]	0.65		1 [v]	-	-	0	合格Passed
其它Other ____					-	-		
包装和标识检验 Packaging & Labelling Checks	13	N/A	无缺陷 No Failure	1			0	合格Passed
* 仅限于Only for E075 & AP E080 ** 仅限于Only for SWCF45 *** CON1/46&CON2/3/4/17&CON69&CON33/66&CON28&CON34&CON9&CON10&CON35/36/37&CON26&CON13&CON14&CON11&CON24&CON27 will be measured as per company's spec and for information, Others will be measured as per ISO4074 measurement. CON1/46 & CON2/3/4/17&CON69&CON33/66&CON28&CON34&CON9&CON10&CON35/36/37&CON26&CON13&CON14&CON11&CON24&CON27 will be measured as per company's spec and for information, Others will be measured as per ISO4074 measurement. **** 仅限于特殊要求. Only for special requirements.								
结论Decision: ☒ 合格Accepted								
报告人Prepared(实验室技术员Lab Technician):	日期Date: 04/07/2024							
批准人Approved By (质控经理QC. MGR.):	日期Date: 04/07/2024							



CS 扫描全能王

3亿人都在用的扫描App