

Certificate of Conformity

Product: Blood Pressure Monitor

Model:	J7136 (BPM HEM-7154-ASH3)	Lot Size:	200
Serial No. Range	20250819401BF	~	20250819600BF
		~	
		~	

Inspection Items		Point of Inspection	Result
Appearance	Exterior	Scratch, Flaw, Burr Recess, Dust, and/ or Contamination	Pass
	Color	Conformity to Boundary Sample	Pass
	Text	Incorrect Texts or Graphics	Pass
	Packaging	Missing Part / Labeling	Pass
Specification	Spec. Drawing	Compliance with Customer's Specification Drawing	Pass
	Pack. Drawing	Compliance with Customer's Packaging Drawing	Pass
Functionality	Function	Malfunction	Pass
	LCD	Incomplete Display (Missing Characters), Illegible or Other Cosmetic Defects	Pass
	Memory	Memory Storage	Pass
Accuracy	Reading	Accuracy of Readings	Pass

This is to certify that the product indicated above has been manufactured in accordance with ISO13485 Medical Device Quality Management System Regulation and meets the specification required by OMRON HEALTHCARE Co.,Ltd. The product has passed the inspection listed in this certificate. This certificate may not be reproduced.

Signature



Satoshi Koike
 Production Quality Improvement Department
 OMRON HEALTHCARE Co., Ltd.

Date (dd/mm/yyyy): 12 / 9 / 2025